

Nabelek-Gorda-YR2 (2013-2014)
LC4x4 Generalized Response and Calibration Factor

These calculations are for the generalized case and assume the signal is in the sensor frequency range giving a flat response. Frequency response ranges are indicated.

SENSOR RESPONSE INFO:

For the Differential Pressure Gauge (DPG) sensitivity:

Using measurements from strain gauge full scale outputs, the average is about 57 mv/7e3 PA. (FSO is 75e-3V/7e3PA). According to Cox *et al* there is a mechanical attenuation factor due to the compressibility of the oil and the compliance of the chamber. They estimate this as 0.86. Willoughby *et al* say they measured the response and deduced this factor as 0.924. Using a value of 0.9 is suggested until a better value can be determined. Thus the STAGE 1 gain should be $5.7e-2 * 0.9/7e3$ or:

S(DPG) = 7.3 μ V/Pa -or- (7.3e-6 V/Pa)

flat response: 0.002 Hz (500 sec) to ~30+ Hz

Frequency response information:

Parameter	Nominal Value	Units
Zeros (1)	0	Rad/s
Poles (1)*	-0.012568	Rad/s
Normalization	1	
Normalization Frequency	0.3	Hz

* *Single Pole @ (1/-79.57).*

For the Trillium-240-OBS seismometer sensitivity:

the manufacturer quotes 1196.5 V*s/m over +/-20V for a full differential signal. SIO-4x4-LP uses only a single-sided input to the A/D, effectively halving the sensitivity, thus:

S(T240-ss) = 598.25 V/m/s

flat response: 0.004167 Hz (240 sec) to 35 Hz

Trillium 240 OBS Seismometer Frequency response information:

(From Trillium 240 OBS User Guide - page 10)

Table 3-2 Poles and zeroes

Parameter	Nominal values	Units
z_n Zeroes	0	rad/s
	0	
	-108	
	-161	
p_n Poles	-0.01815 ±0.01799i	rad/s
	-173	
	-196 ±231i	
	-732 ±1415i	
k Normalization factor	2.316 x 10 ⁹	
S_{sensor} Passband sensitivity at 1Hz	1196.5	V*s/m
f_0 Normalization frequency	1	Hz

**NOTE: Sensor sensitivity listed in Table 3-2 is for full differential response; SIO uses single sided input (halving this number to 598.25 V*s/m).

For the Trillium-40 seismometer sensitivity:

the manufacturer quotes 1553 V*s/m over +/-8V for a full differential signal. SIO-LC4x4 uses only a single-sided input to the A/D, effectively halving the sensitivity, thus:

S(T40-ss) = 776.5 V/m/s

flat response: 0.025 Hz (40 sec) to 50 Hz

Trillium 40 Seismometer Frequency response information:

(From Trillium 40 User Guide - page 40)

Table 9-1 Ground motion response nominal parameters

Symbol	Parameter	Nominal Values	Units
z_n	Zeros	0 0 -68.8 -323 -2530	rad/s
p_n	Poles	-0.1103 ± 0.1110i -86.3 -241 ± 178i -535 ± 719i	rad/s
k	Normalization factor	1.104x 10 ⁵	(rad/s) ²
f_0	Normalization frequency	1	Hz
S	Ground motion sensitivity at f_0	1553	V*s/m

**NOTE: The passband sensitivity listed in Table 4-2 is for full differential response, SIO uses single sided input (halving this number to 776.5 V*s/m).

For the L28LB seismometer sensitivity:

Transduction constant --> 1.57 * sqrt(R-coil) V/m/s with R-coil = 630 ohm nominally this gives 39.53 V/m/s. SIO uses 70% coil current damping, (R-shunt-ss = (7860+51) ohm single-sided, divide by 2 to effective R-shunt damping for differential signal), thus R-shunt-diff = 3956 ohm, which gives:

S(L28) = 34.10 V/m/s

flat response: ~4.5 Hz and above

Frequency response information (for a damped oscillator):

Two zeros at 0, two poles at $\omega_0 \left(\delta \pm i\sqrt{1-\delta^2} \right)$ where $\delta = 0.701$ (damping value).

Parameter	Nominal Value	Units
Zeros (2)	0 0	Rad/s
Poles (2)	19.820 +/- i*20.164	Rad/s
Normalization	-1	
Normalization Frequency	4.5	Hz

ELECTRONICS RESPONSE INFO:

The sensitivity of the A/D is as follows:

With reference filter voltage of $V_{filt} = 100 \text{ ohm}$ the voltage range is $\pm 2.47 \text{ V}$,
 max counts over this range are $-V_{ref} = -6,100,300$ to $+V_{ref} = 6,102,081$.

This gives $S(a/d) = 4.94 / 12,202,381 = 0.405 \times 10^{-6} \text{ V/count} = 0.405 \text{ microV/count}$, or:

$S(a/d) = 0.405 \text{ } \mu\text{V/count}$ -or- $(4.05e-7 \text{ V/count})$

Note: A/D reaches full 24-bit range (i.e. -8388608 to 8388607) @ overvoltage of $\pm 3.3 \text{ V}$. However, the response in this overvoltage range is roughly nonlinear.

Note2: If $V_{filt} = 10 \text{ ohm}$ the voltage range is $\pm 2.50 \text{ V} \rightarrow S(a/d) = 0.410 \text{ microV/count}$.

PRE-AMP GAIN INFO:

Pre-amp gain settings for sensor/channel on all LC4x4 OBS deployments are:

- gain(DPG) = 64**
- gain(T240-ss) = 0.102**
- gain(T40-ss) = 0.200**
- gain(L28) = 64**

Note: To keep the Trillium on scale at the A/D input (max $\pm 2.47 \text{ V}$), signal from the Trillium seismometers output are attenuated using an analog voltage divider on the pre-amp board:

$V-T240-div = R-T240-gnd-eff / (R-T240-sig + R-T240-gnd-eff) = 795 / (6980 + 795) = 0.102$

$V-T40-div = R-T40-gnd-eff / (R-T40-sig + R-T40-gnd-eff) = 1746 / (6980 + 1746) = 0.200$

TOTAL SYSTEM RESPONSE INFO:

Total system response then becomes: $S(\text{total}) = S(a/d) / [S(\text{sensor}) * \text{gain}]$

LC4x4 Generalized Total System Response:

LC4x4-LP units:

DPG pressure response = 0.867 mPa/count (500 sec to ~30+ Hz)
= 8.67e-4 Pa/count

Trillium-240-OBS Velocity response = 6.637 (nm/s)/count (240 sec to 35 Hz)
= 6.637e-9 (m/s)/count

Trillium-40-OBS Velocity response = 2.608 (nm/s)/count (40 sec to 50 Hz)
= 2.608e-9 (m/s)/count

LC4x4-SP units:

DPG pressure response = 0.867 mPa/count (500 sec to ~30+ Hz)
= 8.67e-4 Pa/count

L28 Velocity response = 0.186 (nm/s)/count (~4.5 Hz and above)
= 1.86e-10 (m/s)/count